

Mark Approval Certificate No. 2097/03

Only valid with Terms and Conditions overleaf

RWTÜV Systems GmbH, Postfach 10 32 61, D-45032 Essen

LICENCE HOLDER: REBEC LLC
3511 – 132nd Street SW, Suite No. 100, Lynnwood, WA 98037, USA

MANUFACTURING LOCATION: s. o.

| | | | | |
|------------------------|---------------------|---------------------------|---------------|------------|
| Reference of applicant | Date of application | File reference 20576079 | Date of issue | 21.08.2003 |
| TIM Reber | 02.03.01 / 04.08.03 | 2.4-215/01 Met/Spri / A20 | Valid until | 22.05.2006 |

TEST MARK:

for technical products, components and accessories of technical work equipment



KIND OF DEVICE: Amalgam Separator

Type: RME 2000

Classification: Type 2: Sedimentation system

Maximum waterflow rate: 1 l/min.

TESTED ACCORDING TO: ISO 11143: 1999

ADDITIONAL INFORMATION: REFER TO APPENDIX 1 (TECHNICAL DESCRIPTION FORM)

The test centre for equipment safety of RWTÜV Systems GmbH, designated as a certification body for technical equipment by the Federal Minister for Labour and Social Affairs, together with its associated testing laboratories, hereby confirms:

the type tested device itemized above complies with the state of the art. So far as it is used as a component or accessory for technical equipment it fulfils the specifications as laid down in the current version of the Act on Technical Equipment valid as of 26.08.1992.

Permission to use the Test Mark in accordance with the conditions of contract printed overleaf is hereby granted.

Zertifizierungsstelle für Gerätesicherheit
und Medizinprodukte